



The Regulation and
Quality Improvement
Authority

Announced Inspection

Name of Establishment: Acculase Clinic
Establishment ID No: 10640
Date of Inspection: 14 January 2015
Inspector's Name: Winnie Maguire
Inspection No: 17388

**The Regulation and Quality Improvement Authority
9th floor Riverside Tower, 5 Lanyon Place, Belfast, BT1 3BT
Tel: 028 9051 7500 Fax: 028 9051 7501**

1.0 General Information

Name of establishment:	Acculase Clinic
Address:	c/o Lalapanzi 50 Commercial Road Banbridge BT32 3ES
Telephone number:	07843 062077
Registered organisation/ registered provider:	Ms Caroline Boyd
Registered manager:	Ms Caroline Boyd
Person in charge of the establishment at the time of inspection:	Ms Caroline Boyd
Registration category:	PT(IL)
Date and time of inspection:	14 January 2015 10:00 – 13:00
Date and type of previous inspection:	Approval of new premises variation to registration inspection 9 September 2013
Name of inspector:	Winnie Maguire

2.0 Introduction

The Regulation and Quality Improvement Authority (RQIA) is empowered under The Health and Personal Social Services (Quality, Improvement and Regulation) (Northern Ireland) Order 2003 to inspect independent health care establishments. A minimum of one inspection per year is required.

This is a report of the announced inspection to assess the quality of services being provided. The report details the extent to which the regulations and DHSPPS Minimum Care Standards for Healthcare Establishments, July 2014, measured during the inspection were met.

2.1 Purpose of the Inspection

The purpose of this inspection was to ensure that the service is compliant with relevant regulations, the minimum standards and to consider whether the service provided to clients was in accordance with their assessed needs and preferences. This was achieved through a process of analysis and evaluation of available evidence.

RQIA not only seeks to ensure that compliance with regulations and standards is met but also aims to use inspection to support providers in improving the quality of services. For this reason, inspection involves in-depth examination of an identified number of aspects of service provision.

The aims of the inspection were to examine the policies, procedures, practices and monitoring arrangements for the provision of IPL services, and to determine the provider's compliance with the following:

- The HPSS (Quality, Improvement and Regulation) (Northern Ireland) Order 2003
- The Independent Health Care Regulations (Northern Ireland) 2005
- The Regulation and Improvement Authority (Independent Health Care) (Fees and Frequency of Inspections) (Amendment) Regulations (Northern Ireland) 2011
- The Department of Health, Social Services and Public Safety's (DHSPPS) Minimum Care Standards for Healthcare Establishments

Other published standards which guide best practice may also be referenced during the inspection process.

2.2 Method/Process

Committed to a culture of learning, RQIA has developed an approach which uses self-assessment, a critical tool for learning. The self-assessment was forwarded to the provider prior to the inspection and was reviewed by the inspector prior to the inspection. The inspection process has three key parts; self-assessment, pre-inspection analysis and the visit undertaken by the inspector.

Specific methods/processes used in this inspection include the following:

- Analysis of pre-inspection information and self-assessment
- Discussion with Caroline Boyd, registered provider/manager
- Examination of records
- Tour of the registered premises
- Evaluation and feedback

Any other information received by RQIA about this registered provider and its service delivery has also been considered by the inspector in preparing for this inspection.

The completed self- assessment is appended to this report.

2.3 Consultation Process

During the course of the inspection, the inspector:

Reviewed client feedback questionnaires, issued by the establishment	20
--	----

2.4 Inspection Focus

The inspection sought to establish the level of compliance being achieved with respect to the following DHSPPS Minimum Care Standards for Healthcare Establishments and to assess progress with the issues raised during and since the previous inspection.

- Standard 5 – Patient and Client Partnerships
- Standard 7 – Complaints
- Standard 9 – Clinical Governance
- Standard 16 – Management and Control of Operations
- Standard 48 – Laser and Intense Light Sources

3.0 Profile of Service

Acculase Clinic is located within Lalaparizi beauty salon on the first floor of a commercial building in the centre of Banbridge town.

Private car parking is available for clients.

The establishment is accessible for clients with a disability.

A waiting area and toilet facilities are available for client use.

The establishment's statement of purpose outlines the range of services provided.

The clinic is registered as an independent hospital with the PT (IL) category of registration.

IPL Equipment

Manufacturer:	Lynton
Laser Class:	IPL
Model:	Luminette
Serial Number:	0359-06

Laser Protection Advisor (LPA)

Anna Bass (Lasernet)

Laser Protection Supervisor (LPS)

Caroline Boyd

Authorised User

Caroline Boyd

Type of Treatments Provided

Hair removal

.

4.0 Summary of Inspection

An announced inspection was undertaken by Winnie Maguire on 14 January 2015 from 10:00 to 13:00. The inspector was accompanied by Dr Ian Gillan LPA for RQIA for part of the inspection. The inspection sought to establish the compliance being achieved with respect to The Independent Health Care Regulations (Northern Ireland) 2005, The Regulation and Improvement Authority (Independent Health Care) (Fees and Frequency of Inspections) (Amendment) Regulations (Northern Ireland) 2011, the DHSSPS Minimum Care Standards for Independent Healthcare Establishments and to assess the progress made to address the issues raised during the previous inspection.

One recommendation was made as a result of the previous annual announced inspection on 9 September 2013. The recommendation has been fully addressed.

The inspection focused on the DHSSPS Minimum Care Standards for Independent Healthcare Establishments outlined in section 2.4 of this report.

Caroline Boyd was available during the inspection and for verbal feedback at the conclusion of the inspection.

During the course of the inspection the inspector discussed operational issues, examined a selection of records and carried out a general inspection of the establishment.

The establishment has robust systems in place to obtain the views of clients on a formal and informal basis. The inspector reviewed the completed client feedback questionnaires and found that clients were highly satisfied with the care and treatment provided. Ms Boyd confirmed she will be collating the information from the questionnaires into a summary report which will be made available to clients and other interested parties in an information folder which will be located in the waiting area.

Acculase Clinic has a complaints policy and procedure in place which was found to be in line with the DHSSPS guidance and legislation. No complaints have been received by the establishment, however systems are in place to effectively document, manage and audit complaints. The registered person/manager displayed a good understanding of complaints management.

The registered person/manager is the sole person involved in the service and is therefore responsible for the day to day running of the IPL service and ensuring compliance with the legislation and standards.

The establishment has systems in place to audit the quality of service provided as outlined in the main body of the report.

The inspector also reviewed incident management and found this to be in line with legislation and best practice. No incidents have been recorded by the

establishment however systems are in place to document and manage and report incidents in line with the legislation.

The inspector reviewed the policy and procedures in relation to the absence of the registered manager and whistleblowing. They were found to be in line with legislation and best practice.

A Statement of Purpose and Client Guide were in place which reflected legislative and best practice guidance.

The inspector reviewed the insurance arrangements for the establishment and found that current insurance policies were in place.

Clients are provided with written information regarding the treatment to be provided and fees, risks, complications and expected outcomes are discussed with the client during the initial consultation. Written aftercare instructions are provided following treatment.

There were clear medical treatment protocols and local rules in place. Systems are in place to review the medical treatment protocols and local rules on an annual basis. A service level agreement was in place between the establishment and the LPA.

The establishment has an IPL register which is completed every time the equipment is operated.

Eight client care records were examined and found to be well completed. The records contained the client's personal details, a completed health questionnaire, signed consent form, evidence of patch test and treatment provided.

A risk assessment had been undertaken by the establishment's LPA on 27 November 2014 and all issues identified had been addressed.

Review of the training records confirmed that most mandatory training was up to date and Ms Boyd had received Core of Knowledge training in March 2014. Ms Boyd's training in the safe use and operation of the IPL equipment has been undertaken more than five years ago. A requirement was made to undertake update training in the safe use and operation of the IPL equipment within the next six months. Other staff working in the establishment, but not directly involved in the use of IPL equipment, have received laser safety awareness training.

The IPL service has been moved to a room on the first floor of the building. The LPA for Acculase had visited and updated all the necessary documentation. The fire risk assessment has also been updated. The environment in which the IPL equipment is used was found to be safe and controlled. There were good arrangements for infection prevention and control with suitable hand washing facilities within the room.

Protective eyewear was available for the client and operator as outlined in the local rules.

Laser safety warning signs are displayed when the IPL equipment is in use. The IPL equipment is operated using a key control system. Arrangements are in place for the safe custody of the IPL key when not in use.

Systems were in place to service and maintain the IPL equipment in line with the manufacturers' guidance. The most recent service report was reviewed by the inspector. A laser safety file was in place.

Dr Ian Gillan's report is appended to this report and any action outlined should be addressed.

The certificate of registration was clearly displayed in the reception of the establishment.

One requirement was made as a result of this inspection. This is discussed fully in the main body of the report and in the appended Quality Improvement Plan.

Overall, on the day of inspection, the establishment was found to be providing a safe and effective service.

The inspector would like to extend her gratitude to Ms Boyd for her hospitality and contribution to the inspection process.

5.0 Follow up on Previous Issues

No.	Minimum Standard Ref.	Recommendations	Action taken as confirmed during this inspection	Number of times stated	Inspector's validation of compliance
1	P3	The registered person should ensure training records are available for inspection.	Most records were available for inspection and a certificate of training for basic life support was forwarded to RQIA following the inspection.	One	Compliant

6.0 Inspection Findings

STANDARD 5	
Patient and Client Partnerships:	The views of patients and clients, carers and family members are obtained and acted on in the evaluation of treatment, information and care
<p>Acculase Clinic obtains the views of clients on a formal and informal basis as an integral part of the service they deliver.</p> <p>The inspector reviewed the completed client survey and found that clients were highly satisfied with the treatment provided by Acculase Clinic. No written comments were received from clients.</p> <p>Ms Boyd confirmed the information received from the client survey will be collated into an annual summary report at end January 2015 which will then be made available to clients and other interested parties to read.</p>	

Evidenced by:

Review of client satisfaction surveys
Discussion with staff

STANDARD 7	
Complaints:	All complaints are taken seriously and dealt with appropriately and promptly.
<p>The establishment operates a complaints policy and procedure in accordance with the DHSSPS guidance on complaints handling in regulated establishments and agencies and the legislation. The registered provider/manager demonstrated a good understanding of complaints management.</p> <p>All clients s are provided with a copy of the complaints procedure, which is contained within the Client Guide.</p> <p>The inspector reviewed the complaints register and found that no complaints had been received by the establishment; however systems are in place to effectively document and manage complaints.</p>	

Evidenced by:

Review of complaints procedure

Complaint procedure made available to clients and other interested parties

Discussion with staff

Review of complaints records

STANDARD 9	
Clinical Governance:	Patients and clients are provided with safe and effective treatment and care based on best practice guidance, demonstrated by procedures for recording and audit.
<p>The registered provider/manager ensures the establishment delivers a safe and effective service in line with the legislation, other professional guidance and minimum standards.</p> <p>The establishment has systems in place to audit the quality of service provided. The inspector reviewed the following audits as part of the inspection process:</p> <ul style="list-style-type: none"> • Client survey • Reflective practice <p>The establishment has an incident policy and procedure in place which includes reporting arrangements to RQIA. No incidents have occurred within the establishment since registration; however systems are in place to document and manage incidents appropriately.</p>	

Evidenced by:

- Review of policies and procedures**
- Discussion with registered provider/manager**
- Review of audits**
- Review of incident management**

STANDARD 16	
Management and Control of Operations:	Management systems and arrangements are in place that ensure the delivery of quality treatment and care.
<p>Ms Boyd is the sole person involved in the operation of the IPL service.</p> <p>The establishment has a policy and procedure in place to ensure that RQIA is notified if the registered manager is absence for more than 28 days.</p> <p>The inspector reviewed the establishment’s Client Guide and Statement of Purpose and found them to be in line with the legislation.</p> <p>There is a written policy on "Whistle Blowing" and written procedures that identify to whom staff report concerns about poor practice and the support mechanisms available to those staff.</p> <p>The inspector discussed the insurance arrangements within the establishment and confirmed current insurance policies were in place. The certificate of registration was clearly displayed in the reception area of the premises.</p>	

Evidenced by:

- Review of policies and procedures**
- Review of Client Guide**
- Review of Statement of Purpose**
- Review of insurance arrangements**

STANDARD 48**Laser and Intense Light Sources:**

Laser and intense light source procedures are carried out by appropriately trained staff in accordance with best practice.

Clients are provided with an initial consultation to discuss their treatment and any concerns they may have. The establishment has a list of fees available for each IPL procedure. Fees for treatments are agreed during the initial consultation and may vary depending on the treatment provided and the individual requirements of the client.

Clients are asked to complete a health questionnaire. There are systems in place to contact the medical support service, with their consent, for further information and guidance if necessary.

Clients are provided with written information on the specific IPL procedure to be provided that explains the risks, complications and expected outcomes of the treatment.

IPL procedures are carried out by a trained operator in accordance with medical treatment protocols produced by Dr Paul Myers in March 2014. Systems are in place to review the medical treatment protocols on an annual basis.

The medical treatment protocols set out:

- Indications
- Contraindications
- Technique
- Pre-treatment tests
- Pre-treatment care
- Post-treatment care
- Recognition of treatment related problems
- Procedure if anything goes wrong with the treatment
- Permitted variation on machine variables
- Procedure in the event of equipment failure

There was written confirmation of the appointment and duties of a certified LPA which is reviewed on an annual basis. The inspector reviewed the service level agreement between the establishment and the LPA which expires in April 2015.

The establishment has local rules in place which have been developed by their LPA on 27 November 2014.

The local rules cover:

- The potential hazards associated with lasers
- Controlled and safe access
- Authorised operator's responsibilities
- Methods of safe working

- Safety checks
- Personal protective equipment
- Prevention of use by unauthorised persons
- Adverse incidents procedures

The laser protection supervisor has overall responsibility for safety during IPL treatments as recorded within the local rules.

A list of authorised users is maintained and the authorised user has signed to state that they have read and understood the local rules and medical treatment protocols.

The establishment has an IPL register which is completed every time the equipment is operated and includes:

- The name of the person treated
- The date
- The operator
- The treatment given
- The precise exposure
- Any accident or adverse incident

Eight client care records were reviewed and found to contain information regarding the client's personal details, a completed health questionnaire, signed consent form, record of treatment and evidence of a patch test being undertaken. The inspector noted records of contacts with the medical support service had been fully recorded in clients care records outlining advice sought and advice given.

Clients are provided with written aftercare instructions following treatment.

The establishment's LPA completed a risk assessment of the premises in November 2014 and all recommendations made by the LPA had been addressed.

Ms Boyd had completed training in Core of Knowledge in March 2014. On review of training records it was noted it had been over five years since Ms Boyd had received training on the safe use and application of the IPL equipment. A requirement was made to undertake update training within the next six months.

Review of the training records confirmed that the authorised user had also undertaken the following required mandatory training in line with RQIA guidance:

- Basic life support annually (certificate forwarded to the inspector following inspection.)
- Fire safety annually
- Infection prevention and control annually

All other staff employed at the establishment, but not directly involved in the use of the IPL equipment, had received laser safety awareness training.

The environment in which the IPL equipment is used was found to be safe and

controlled to protect other persons while treatment is in progress. The controlled area is clearly defined and not used for other purposes, or as access to areas, when treatment is being carried out.

The fire risk assessment has been updated following the IPL service moving to a room on the first floor of the building.

There were good arrangements for infection prevention and control with suitable hand washing facilities within the room.

When the IPL equipment is in use, the safety of all persons in the controlled area is the responsibility of the LPS.

Laser safety warning signs are displayed when the IPL equipment is in use and removed when not in use, as described within the local rules.

Protective eyewear is available for the client and operator as outlined in the local rules. The inspector reviewed the protective eyewear available as part of the inspection process.

The door to the treatment room is locked when the IPL equipment is in use but can be opened from the outside in the event of an emergency.

The IPL equipment is operated using a key. Arrangements are in place for the safe custody of the IPL key when not in use.

There are arrangements in place to service and maintain the IPL equipment in line with the manufacturer's guidance. The most recent service report dated September 2014 was reviewed as part of the inspection process.

There is a laser safety file in place that contains all of the relevant information relating to the IPL equipment.

Evidenced by:

Discussion with authorised user
Review of policies and procedures
Review of information provided to clients
Review of local rules
Review of medical treatment protocols
Review of IPL register
Review of client care records
Review of LPA's risk assessment
Review of training records
Review of premises and controlled area
Review of maintenance records
Review of laser safety file

7.0 Quality Improvement Plan

The details of the Quality Improvement Plan appended to this report were discussed with Ms Caroline Boyd as part of the inspection process.

The timescales for completion commence from the date of inspection.

The registered provider / manager is required to record comments on the Quality Improvement Plan.

Matters to be addressed as a result of this inspection are set in the context of the current registration of your premises. The registration is not transferable so that in the event of any future application to alter, extend or to sell the premises the RQIA would apply standards current at the time of that application.

Enquiries relating to this report should be addressed to:

Winnie Maguire
The Regulation and Quality Improvement Authority
9th Floor
Riverside Tower
5 Lanyon Place
Belfast
BT1 3BT

Winnie Maguire
Inspector / Quality Reviewer

Date



The Regulation and
Quality Improvement
Authority

Quality Improvement Plan
Announced Inspection
Acculase Clinic
14 January 2015



The areas where the service needs to improve, as identified during this inspection visit, are detailed in the inspection report and Quality Improvement Plan.

The specific actions set out in the Quality Improvement Plan were discussed with Ms Caroline Boyd either during or after the inspection visit.

Any matters that require completion within 28 days of the inspection visit have also been set out in separate correspondence to the registered persons.

Registered providers / managers should note that failure to comply with regulations may lead to further enforcement and/or prosecution action as set out in The HPSS (Quality, Improvement and Regulation) (Northern Ireland) Order 2003.

It is the responsibility of the registered provider / manager to ensure that all requirements and recommendations contained within the Quality Improvement Plan are addressed within the specified timescales.

Matters to be addressed as a result of this inspection are set in the context of the current registration of your premises. The registration is not transferable so that in the event of any future application to alter, extend or to sell the premises the RQIA would apply standards current at the time of that application.

STATUTORY REQUIREMENT

This section outlines the actions which must be taken so that the registered person/s meets legislative requirements based on The HPSS (Quality, Improvement and Regulation) (Northern Ireland) Order 2003, The Independent Health Care Regulations (Northern Ireland) 2005 and The Regulation and Improvement Authority (Independent Health Care) (Fees and Frequency of Inspections) (Amendment) Regulations (Northern Ireland) 2011.

NO.	REGULATION REFERENCE	REQUIREMENT	NUMBER OF TIMES STATED	DETAILS OF ACTION TAKEN BY REGISTERED PERSON(S)	TIMESCALE
1	18(2)	The registered person must ensure they undertake update training on the safe use and application of the IPL equipment. Ref: Standard 48	One	<i>Booked update Clinical Training + Flights For Monday 11th May 2015 So will be completed then.</i>	Six months

The registered provider/manager is required to detail the action taken, or to be taken, in response to the issue(s) raised in the Quality Improvement Plan. The Quality Improvement Plan is then to be signed below by the registered provider and registered manager and returned to:

Winnie Maguire
 The Regulation and Quality Improvement Authority
 9th floor
 Riverside Tower
 5 Lanyon Place
 Belfast
 BT1 3BT

SIGNED: Caroline Boyd

NAME: CAROLINE Boyd
 Registered Provider

DATE: 24th February 2015

SIGNED: Caroline Boyd

NAME: CAROLINE Boyd
 Registered Manager

DATE: 24th Feb 2015

QIP Position Based on Comments from Registered Persons		Yes	No	Inspector	Date
A	Quality Improvement Plan response assessed by inspector as acceptable	✓		W Maguire	3/3/15
B	Further information requested from provider				



**The Regulation and
Quality Improvement
Authority**

REGULATION AND QUALITY

10 OCT 2014

IMPROVEMENT AUTHORITY



Pre-Inspection Self-Assessment Laser/IPL Service

Name of Establishment: Acculase Clinic
Establishment ID No: 10640
Date of Inspection: 7 October 2014
Inspector's Name: Winnie Maguire
Inspection No: 17388

THE REGULATION AND QUALITY IMPROVEMENT AUTHORITY
9th floor Riverside Tower, 5 Lanyon Place, Belfast, BT1 3BT
Tel: 028 9051 7500 Fax: 028 9051 7501

1.0 Introduction

The Regulation and Quality Improvement Authority (RQIA) is empowered under The Health and Personal Social Services (Quality, Improvement and Regulation) (Northern Ireland) Order 2003 to inspect independent health care establishments. A minimum of one inspection per year is required.

The aim of inspection is to examine the policies, procedures, practices and monitoring arrangements for the provision of laser/IPL services, and to determine the provider's compliance with the following:

- The HPSS (Quality, Improvement and Regulation) (Northern Ireland) Order 2003
- The Independent Health Care Regulations (Northern Ireland) 2005
- The Regulation and Improvement Authority (Independent Health Care) (Fees and Frequency of Inspections) (Amendment) Regulations (Northern Ireland) 2011
- The Department of Health, Social Services and Public Safety's (DHSSPS) Minimum Care Standards for Independent Healthcare Establishments July 2014

Other published standards which guide best practice may also be referenced during the inspection process.

2.0 Self-Assessment

Committed to a culture of learning, RQIA has developed an approach which uses self-assessment.

Where asked in the self-assessment you are required to indicate a yes or no response. You are also asked to provide a brief narrative in the "text box" where applicable.

Following completion of the self-assessment, please return to RQIA by the date specified.

The self-assessment will be appended to the report and made available to the public. No amendments will be made by RQIA to your self-assessment response.

3.0 Self-Assessment Tool

Management of Operations

	YES	NO
Has any structural change been made to the premises since the previous inspection?	✓	
Have any changes been made to the management structure of the establishment since the previous inspection?		✓
Yes, please comment <i>I have already advised of structural changes + inspection date has been put back</i>		

Policies and Procedures

	YES	NO
Does the establishment have a policy and procedure manual in place which is reviewed at least every 3 years or as changes occur?	✓	
Are the policies and procedures for all operational areas in line with legislation and best practice guidelines?	✓	
Do all policies and procedures contain the date of issue, date of review and version control?	✓	
Are all policies and procedures ratified by the registered person?	✓	
No, please comment		

Records Management

	YES	NO
Does the establishment have a policy and procedure in place for the creation, storage, transfer, retention and disposal of and access to records in line with the legislation?	✓	
Are care records maintained for each individual client?	✓	
Are arrangements in place to securely store client care records?	✓	
No, please comment		

Patient Partnerships

	YES	NO
Does the establishment have systems in place to obtain the views of clients regarding the quality of treatment, care and information provided?	✓	
Does the establishment make available a summary report of client feedback to clients and other interested parties?	✓	
No, please comment		

Medical Emergencies

	YES	NO
Are arrangements in place to deal with medical emergencies?	✓	
No, please comment		

Complaints

	YES	NO
Does the establishment have a complaints policy and procedure in place which is in line with the legislation and the DHSSPS guidance on complaints handling in regulated establishments and agencies April 2009?	✓	
Are all complaints documented, fully investigated and have outcomes recorded in line with the legislation and the establishment's complaints policy and procedure?	✓	
No, please comment		

Incidents

	YES	NO
Does the establishment have an incident policy and procedure in place which complies with the legislation and RQIA guidance?	✓	
Are all incidents reported, documented, fully investigated and have outcomes recorded in line the legislation, RQIA guidance and the establishment's policy and procedure?	✓	
No, please comment		

Infection Prevention and Control

	YES	NO
Does the establishment have an infection prevention and control policy and procedure in place?	✓	
Are appropriate arrangements in place to decontaminate equipment between clients?	✓	
No, please comment		

Recruitment of staff

	YES	NO
Does the establishment have a recruitment and selection policy and procedure in place?		✓
Is all information outlined in Schedule 2 of the Independent Health Care Regulations (Northern Ireland) 2005 retained and available for inspection?	✓	
Have all authorised users (recruited since registration with RQIA) had an enhanced AccessNI disclosure undertaken, prior to commencing employment?	✓	
No, please comment <i>There will be no recruitment in the foreseeable future.</i>		

Mandatory Training

	YES	NO
Are arrangements in place for all new authorised users to participate in an induction programme?	✓	
Are training records available which confirm that the following mandatory training has been undertaken:		
AUTHORISED USERS	YES	NO
Core of knowledge training – within the past 5 year years	✓	
Application training for all equipment and all laser/IPL treatments provided - within the past 5 years		
Infection prevention and control training – annually	✓	
Fire safety – annually	✓	
Basic life support – annually or valid certificate e.g. First Aid at Work which is valid for 3 years	✓	
OTHER STAFF – NOT INVOLVED IN LASER/IPL SERVICES (If applicable)	YES	NO
Laser safety awareness training – annually	✓	
If No, please comment		

Appraisal

	YES	NO
Does the establishment have an appraisal policy and procedure in place?	✓	
Are systems in place to provide recorded annual appraisals for authorised users? (if applicable)	✓	
No, please comment		

Qualifications of Medical Practitioners and Nurses

	YES	NO
Are systems in place to ensure medical and nursing staff, who are authorised users, have a current registration with their relevant professional bodies?	✓	
Are systems in place to ensure that medical practitioners have up to date professional indemnity insurance?	✓	✓
Are systems in place to ensure that medical practitioners have an annual appraisal undertaken with a trained medical appraiser?		✓
Are arrangements in place to ensure medical practitioners have a responsible officer?		✓
No, please comment		

Lasers/IPL Service

	YES	NO
Does the establishment have a certified Laser Protection Advisor (LPA)?	✓	
Has the establishment an up to date LPA report?	✓	
Has the establishment an up to date risk assessment undertaken by their LPA?	✓	
Does the establishment have up to date local rules in place?	✓	
Does the establishment have up to date medical treatment protocols in place?	✓	
Are systems in place to review local rules and medical treatment protocols on an annual basis?	✓	
Does the establishment have arrangements in place for a medical support service?	✓	
Does the establishment have a list of authorised users?	✓	
Does the establishment have arrangements in place for maintenance and servicing of equipment in line with manufacturer's guidance?	✓	
Does the establishment have protective eyewear in place, as outlined in the local rules?	✓	
Is the controlled area clearly defined?	✓	
Is the door to the treatment room where the laser/IPL used locked during treatment and can be opened from the outside in the event of an emergency?	✓	
Does the establishment display laser/IPL warning signs as outlined in the local rules?	✓	
Are arrangements in place for the safe custody of laser/IPL keys and/or keypad codes?	✓	
Does the establishment have a laser/IPL safety file in place?	✓	

Does the establishment have a laser/IPL register(s) in place?	✓	
<p>No, please comment</p>		

4.0 Declaration

To be signed by the registered provider or registered manager for the establishment.

I hereby confirm that the information provided above is, to the best of my knowledge, accurately completed.

Name	Signature	Designation	Date
<i>Carole Boyd</i>	<i>Carole Boyd</i>	Registered owner MANAGER	8 Oct 14