

Inspection Report

23 July 2025



Laserase Medical Aesthetics

Type of service: Independent Hospital - Cosmetic Laser /
Intense Pulsed Light / Private Doctor
Address: 525 Antrim Road, Belfast, BT15 3BS
Telephone number: 028 9077 7772

www.rqia.org.uk

Assurance, Challenge and Improvement in Health and Social Care

Information on legislation and standards underpinning inspections can be found on our website <https://www.rqia.org.uk/> [The Independent Health Care Regulations \(Northern Ireland\) 2005](#) and [Minimum Care Standards for Independent Healthcare Establishments \(July 2014\)](#)

1.0 Service information

<p>Organisation/Provider: LMS (NI) Limited</p> <p>Responsible Individual Mrs Therese McKeown</p>	<p>Registered Manager: Mr Rory McKeown</p> <p>Date registered: 24 January 2024</p>
<p>Person in charge at the time of inspection: Mr Rory McKeown</p>	
<p>Categories of care: Independent Hospital (IH) Prescribed techniques or prescribed technology: establishments using Class 3B or Class 4 lasers PT(L) and Prescribed techniques or prescribed technology: establishments using intense light sources PT(IL) and Private Doctor (PD)</p>	
<p>Brief description of how the service operates: Laserase Medical Aesthetics provides a range of cosmetic/aesthetic treatments. This inspection focused solely on those treatments using a Class 4 laser, an intense pulse light (IPL) machine and the private doctor service that fall within regulated activity and the categories of care for which the establishment is registered with RQIA.</p> <p>The registered manager confirmed that private doctor services are not currently provided in the establishment.</p> <p>Equipment available in the service:</p> <p>Laser equipment:</p> <p>Manufacturer: Euphotonics Model: NLite – V Multifunction Laser Serial Number: NL3429 Laser Class: Class 4 Wavelength: 580nm – 590nm</p> <p>Manufacturer: Candela Model: Picoway Serial Number: 1064/532/785 Laser Class: Class 4 Wavelength: 1064/532/785nm</p>	

Multi-platform equipment:

Manufacturer: Lumenis
 Model: M22tm- Combined IPL & Yag Q Switch
 Serial Number: 21046
 Laser Class: Class 4
 Wavelength: IPL 400 – 1200 Nd / YAG 1064nm / ResurFX 1565nm
 Hand pieces: 3

The Lumenis M22 is a multi-platform machine that is capable of operating as a laser and an IPL by changing the hand piece.

Types of laser treatments provided:

Vascular hair removal, tattoo removal, acne, acne scarring, wrinkle reduction

Types of IPL treatments provided:

Hair removal, skin rejuvenation, red veins, acne, pigmented skin, red flushing skin

2.0 Inspection summary

This was an announced inspection, undertaken by a care inspector on 23 July 2025 from 9.30 am to 11.30 pm.

The purpose of the inspection was to assess progress with areas for improvement identified during the last care inspection and to assess compliance with the legislation and minimum standards.

There was evidence of good practice concerning staff recruitment; authorised operator training; safeguarding; laser and IPL safety; management of medical emergencies; infection prevention and control (IPC); the management of clinical records; and effective communication between clients and staff.

Additional areas of good practice identified included maintaining client confidentiality, ensuring the core values of privacy and dignity were upheld and providing the relevant information to allow clients to make informed choices.

No immediate concerns were identified regarding the delivery of front line client care.

3.0 How we inspect

RQIA is required to inspect registered services in accordance with legislation. To do this, we gather and review the information we hold about the service, examine a variety of relevant records, meet and talk with staff and management and observe practices on the day of the inspection.

The information obtained is then considered before a determination is made on whether the establishment is operating in accordance with the relevant legislation and minimum standards. Examples of good practice are acknowledged and any areas for improvement are discussed with the person in charge and detailed in the quality improvement plan (QIP).

4.0 What people told us about the service

Posters were issued to Laserase Medical Aesthetics by RQIA prior to the inspection inviting clients and staff to complete an electronic questionnaire.

No clients submitted responses.

Two staff submitted questionnaire responses. Staff responses indicated that they felt client care was safe, effective, that clients were treated with compassion and that the service was well led. All staff indicated that they were satisfied or very satisfied with each of these areas of client care. None of the staff responses included comments.

5.0 The inspection

5.1 What has this service done to meet any areas for improvement identified at or since last inspection?

The last inspection to Laserase Medical Aesthetics was undertaken on 21 June 2024; no areas for improvement were identified.

5.2 Inspection outcome

5.2.1 How does this service ensure that staffing levels are safe to meet the needs of clients and staff are suitably trained?

Mr McKeown told us there are sufficient staff in the various roles to fulfil the needs of the establishment and clients.

Mr McKeown confirmed that laser and IPL treatments are only carried out by authorised operators. A register of authorised operators for the laser and IPL equipment is maintained and kept up to date.

Mr McKeown confirmed that an induction programme and training will be provided to newly recruited authorised operators on commencement of employment.

A review of training records evidenced that authorised operators have up to date training in core of knowledge training, application training for the equipment in use, basic life support, IPC, fire safety awareness and safeguarding adults at risk of harm in keeping with the RQIA training guidance.

All other staff employed at the establishment, but not directly involved in the use of the laser or IPL equipment, had received laser safety awareness training.

Evidence was available that staff who have professional registration undertake continuing professional development (CPD) in accordance with their professional body's recommendations. Discussion with Mr McKeown and review of documentation confirmed that authorised users take part in appraisal on an annual basis.

It was determined that appropriate staffing levels were in place to meet the needs of clients and that staff are suitably trained.

5.2.2 How does the service ensure that recruitment and selection procedures are safe?

Recruitment and selection policies and procedures were in place, which adhered to legislation and best practice guidance for the recruitment of authorised operators.

There have been no authorised operators recruited since the previous inspection. During discussion, Mr McKeown confirmed that should authorised operators be recruited in the future, all recruitment documentation as outlined in Schedule 2 of The Independent Health Care Regulations (Northern Ireland) 2005 would be sought and retained for inspection.

Discussion with Mr McKeown confirmed that he had a clear understanding of the legislation and best practice guidance in relation to recruitment and selection.

It was determined that the recruitment of authorised operators complies with the legislation and best practice guidance.

5.2.3 How does the service ensure that it is equipped to manage a safeguarding issue should it arise?

Mr McKeown stated that laser and IPL treatments are not provided to persons under the age of 18 years.

A policy and procedure was in place for the safeguarding and protection of adults and children at risk of harm. The policy included the types and indicators of abuse and distinct referral pathways in the event of a safeguarding issue arising with an adult. The relevant contact details were included for onward referral to the local Health and Social Care Trust should a safeguarding issue arise.

Discussion with Mr McKeown confirmed that he was aware of the types and indicators of abuse and the actions to be taken in the event of a safeguarding issue being identified.

Review of records demonstrated that Mr McKeown, as the safeguarding lead, has completed formal level two training in safeguarding adults in keeping with the Northern Ireland Adult Safeguarding Partnership (NIASP) training strategy (revised 2016) and minimum standards.

It was confirmed that a copy of the regional guidance document entitled Adult Safeguarding Prevention and Protection in Partnership (July 2015) was available for reference.

It was determined that the service had appropriate arrangements in place to manage a safeguarding issue should it arise.

5.2.4 How does the service ensure that medical emergency procedures are safe?

All authorised operators had up to date training in basic life support and Mr McKeown was aware of what action to take in the event of a medical emergency.

There was a written protocol in place for dealing with recognised medical emergencies.

It was determined that the service had appropriate arrangements in place to manage a medical emergency.

5.2.5 How does the service ensure that it adheres to infection prevention and control (IPC) and decontamination procedures?

The IPC arrangements were reviewed throughout the establishment, to evidence that the risk of infection transmission to clients, visitors and staff was minimised.

There was an overarching IPC policy and associated procedures in place. A review of these documents demonstrated that they were comprehensive and reflected legislation and best practice guidance.

Two treatment rooms are used for laser and IPL treatments. The treatment room viewed during the inspection was found to be clean and clutter free. Discussion with Mr McKeown evidenced that appropriate procedures were in place for the decontamination of equipment between use.

Hand washing facilities were available and adequate supplies of personal protective equipment (PPE) were provided.

Cleaning schedules for the establishment were in place. As discussed previously, authorised operators had up to date training in IPC.

Mr McKeown is aware that the Department of Health (DOH) and Public Health Agency (PHA) websites provide advisory information, guidance and alerts with regards to IPC.

It was determined that the service had appropriate arrangements in place in relation to IPC and decontamination.

5.2.6 How does the service ensure the environment is safe?

The service has two laser and IPL treatment rooms. The premises were maintained to a good standard of maintenance and décor.

Observations made evidenced that a carbon dioxide (CO₂) fire extinguisher is available which has been serviced within the last year. It was confirmed that the fire risk assessment is reviewed annually.

It was determined that appropriate arrangements were in place to maintain the environment.

5.2.7 How does the service ensure that laser and IPL procedures are safe?

A laser safety file was in place which contained the relevant information in relation to laser and IPL equipment. There was written confirmation of the appointment and duties of a certified laser protection advisor (LPA) which is reviewed on an annual basis. The service level agreement between the establishment and the LPA was reviewed and this expires during December 2025.

Up to date, local rules were in place which have been developed by the LPA. The local rules contained the relevant information about the laser and IPL equipment being used.

The establishment's LPA last completed a risk assessment of the premises during December 2023 and all recommendations made by the LPA have been addressed.

Mr McKeown confirmed that laser and IPL procedures are carried out following medical treatment protocols. The medical treatment protocols had been produced by a named registered medical practitioner.

It was demonstrated that the protocols contained the relevant information about the treatments being provided with the exception of one protocol associated with the M22 ResurFX platform. This was brought to the attention of Mr McKeown, who agreed to contact his Medical Expert Practitioner. Following the inspection, evidence was submitted to RQIA to confirm that the protocol had been sought and retained on file and signed by both AOs.

It was confirmed that medical treatment protocols are due to expire during December 2025. It was established that systems are in place to review the medical treatment protocols when due.

Mr McKeown, as the laser protection supervisor (LPS), has overall responsibility for safety during laser and IPL treatments and a list of authorised operators is maintained. Authorised operators had signed to state that they had read and understood the local rules and medical treatment protocols.

When the laser or IPL equipment is in use, the safety of all persons in the controlled area is the responsibility of the LPS.

The environment in which the laser and IPL equipment is used was found to be safe and controlled to protect other persons while treatment is in progress. The controlled area is clearly defined and not used for other purposes, or as access to areas, when treatment is being carried out.

The door to each treatment room is locked when the laser or IPL equipment is in use but can be opened from the outside in the event of an emergency. Authorised operators were aware that the laser safety warning signs should only be displayed when the equipment is in use and removed when not in use.

The N-Lite and Picoway laser machines are operated using a key and the Lumenis M22 multiplatform is operated using a key code. Arrangements are in place for the safe custody of the key and keypad code when not in use.

Protective eyewear is available for the client and operator as outlined in the local rules.

Laserase Medical Aesthetics has four registers, one for each machine and one for laser tattoo removal.

Mr McKeown told us authorised operators complete the relevant section of the registers every time the equipment is operated.

The four registers were reviewed and found to include:

- the name of the person treated
- the date
- the operator
- the treatment given
- the precise exposure
- any accident or adverse incident

There are arrangements in place to service and maintain the laser and IPL equipment in line with the manufacturer's guidance. The most recent service reports for the laser and IPL equipment were reviewed.

As a result of the action taken following inspection, it was determined that appropriate arrangements were in place to operate the laser and IPL equipment.

5.2.8 How does the service ensure that clients have a planned programme of care and have sufficient information to consent to treatment?

Mr McKeown confirmed that clients are provided with an initial consultation to discuss their treatment and any concerns they may have. There is written information for clients that provides a clear explanation of any treatment and includes pre and post treatment information.

The service has a list of fees available for each laser and IPL procedure. Fees for treatments are agreed during the initial consultation and may vary depending on the type of treatment provided and the individual requirements of the client.

During the initial consultation each client's personal information is recorded, and clients are asked to complete a health questionnaire.

A sample of client care records were reviewed. There was an accurate and up to date electronic treatment record for every client which included:

- client details
- medical history
- signed consent form
- general practitioner (GP) details
- skin assessment (where appropriate)
- patch test (where appropriate)
- record of treatment delivered including number of shots and fluence settings (where appropriate)

Observations made evidenced that client records are securely stored. A policy and procedure was available which included the creation, storage, recording, retention and disposal of records and data protection.

The service has a policy for advertising and marketing.

As a result of the action taken following inspection, it was determined that appropriate arrangements are in place to ensure that clients have a planned programme of care and have sufficient information to consent to treatment.

5.2.9 How does the service ensure that clients are treated with dignity, respect and are involved in the decision making process?

Discussion with Mr McKeown regarding the consultation and treatment process confirmed that clients are treated with dignity and respect. The consultation and treatment are provided in a private room with the client and authorised operator present. Information is provided to the client in verbal and written form at the initial consultation and subsequent treatment sessions to allow the client to make choices about their care and treatment and provide informed consent.

Mr McKeown told us that clients are provided with the opportunity to complete a satisfaction survey when their treatment is complete. The results of these are collated to provide an anonymised summary report which is made available to clients and other interested parties. Mr McKeown confirmed that an action plan would be developed to inform and improve services provided, if appropriate.

Review of the most recent client satisfaction report, dated January 2025, found that clients were very satisfied with the quality of treatment, information and care received.

It was determined that appropriate arrangements were in place to ensure that clients are treated with dignity, respect and are involved in decisions regarding their choice of treatment.

5.2.10 How does the registered provider assure themselves of the quality of the services provided?

Where the entity operating the service is a corporate body or partnership or an individual owner who is not in day to day management of the service, unannounced quality monitoring visits by the registered provider must be undertaken and documented every six months; as required by Regulation 26 of The Independent Health Care Regulations (Northern Ireland) 2005.

Mr McKeown, as the registered manager, is the nominated individual with overall responsibility for the day-to-day management of the establishment. He is responsible for reporting to Mrs Therese McKeown, Responsible individual.

At the last inspection we we advised that, in order for Mrs McKeown to monitor the quality of services provided, unannounced monitoring visits to the premises were required at least every six months in accordance with legislation. Reports of these visits along with any identified actions should be made available for inspection.

Mr McKeown confirmed that these monitoring visits had not yet commenced as previously advised. Mr McKeown agreed to address this matter as priority with Mrs McKeown. Following inspection, confirmation was received that a Regulation 26 visit had been undertaken by Mrs McKeown on 31 July 2025 and the subsequent report was made available to RQIA.

Policies and procedures were available outlining the arrangements associated with the laser and IPL treatments. Observations made confirmed that policies and procedures were indexed, dated and systematically reviewed on a three yearly basis or more frequently if required.

The arrangements for the management of complaints and incidents were reviewed to ensure that they were being managed in keeping with legislation and best practice guidance.

The complaints policy and procedure provided clear instructions for clients and staff to follow. Clients were made aware of how to make a complaint by way of the client's guide.

Arrangements were in place to record any complaint received in a complaints register and retain all relevant records including details of any investigation undertaken, all communication with complainants, the outcome of the complaint and the complainant's level of satisfaction.

Mr McKeown confirmed that no complaints had been received since the previous inspection.

Discussion with Mr McKeown confirmed that an incident policy and procedure was in place which includes the reporting arrangements to RQIA. Mr McKeown confirmed that incidents would be effectively documented and investigated in line with legislation and reported to RQIA and other relevant organisations in accordance with legislation and RQIA [Statutory Notification of Incidents and Deaths](#). Arrangements are in place to audit adverse incidents to identify trends and improve service provided.

Mr McKeown demonstrated an understanding of her role and responsibility in accordance with legislation.

Mr McKeown confirmed that the statement of purpose and client's guide are kept under review, revised and updated when necessary and available on request.

The RQIA certificate of registration was displayed in a prominent place.

Observation of insurance documentation confirmed that current insurance policies were in place.

As a result of the action taken following inspection, it is determined that suitable arrangements are in place to enable the responsible individual to assure themselves of the quality of the services provided.

5.2.11 Does the service have suitable arrangements in place to record equality data?

The arrangements in place in relation to the equality of opportunity for clients and the importance of staff being aware of equality legislation and recognising and responding to the diverse needs of clients was discussed with Mr McKeown.

6.0 Quality Improvement Plan/Areas for Improvement

This inspection resulted in no areas for improvement being identified. Findings of the inspection were discussed with Mr McKeown, Registered Manager, as part of the inspection process and can be found in the main body of the report.



The **Regulation** and
Quality Improvement
Authority

The Regulation and Quality Improvement Authority
James House
2-4 Cromac Avenue
Gasworks
Belfast
BT7 2JA

Tel 028 9536 1111
Email info@rqia.org.uk
Web www.rqia.org.uk
 [@RQIANews](https://twitter.com/RQIANews)

Assurance, Challenge and Improvement in Health and Social Care